

IC10/20

Dual Incubator

For Biological Indicators



Usage

Incubator for biological indicators at 37 ± 2 °C or 60 ± 2 °C.

Applicable Regulation

European Union Low Voltage Directive 2006 / 95/ EC.
Electromagnetic Compatibility Directive 2004/108/CE.
RoHS Directive 2011/65/UE.

Product Classification

Class 1, according to risk.

Authorization

Designed under Quality Management System standards ISO 13485:2003/NS-EN ISO 13485:2012.

Characteristics

Dimensions: 12.0 cm high, 12.3 cm diameter.

14 metal positions (1.3 cm diameter, 3.0 cm deep) to incubate BT23, culture media and protein pen.

32 metal positions (0.9 cm diameter, 3.0 cm deep) to incubate BT10, BT20, BT30, BT80, BT90, BT91, BT100, BT101, BT200, BT202, BT21, BT22, BT24.

Special hole for external thermometer.

1 ampoule crusher.

Weight: 0.9 kgs.

Voltage range: 100 – 240 V

Power: 20 W

Frequency: 50 - 60 Hz

Environmental conditions during manufacture

Room temperature 15-35 °C, RH 35-85 %

Storage conditions

T = 10-30 °C, RH 30-80 %, keep in a dark place.

Transportation conditions

Storage conditions should be strictly followed. The product should be transported in closed and reinforced boxes in order to avoid damages. Product transportation does not represent any risk for human health.

Shelf-life

10 years

Packing

One unit per box.

Packing information: product code and description, and manufacturer information.

Labelling

Same label on product and product's box: product code and description, batch number, serial number and manufacturer information.

Possible target markets

Healthcare, Food, Pharmaceutical and medical devices industries.

Other important information (optional)

Incubator IC10/20 has a switch button located at the base of the incubator to select the temperature at 37 ± 2 °C or 60 ± 2 °C.

Note: When necessary, limits of each product's parameters should be specified.

Description of identified and/or predictable risks which could represent a risk for the product

Do not pour any liquid inside.

Do not immerse into any liquid.

Only for indoor use.

Disconnect the power cord before cleaning.

Do not use abrasive or corrosive cleaners or disinfectants.

Make sure that the incubator is connected to a properly rated power cord.